

# Model ED-S 56 | Drying and heating chambers with natural convection

The new products in the Solid.Line impress with their tried-and-tested BINDER quality and reliability. They are suited to drying and heating. The attributes of the Solid.Line are geared towards the application in question and therefore used in research and quality assurance.

## BENEFITS

- Easy handling
- Accurate temperature control
- Simple and ergonomic door opening



Model 56

## IMPORTANT FEATURES

- Temperature range: +7°C above ambient temperature to +250 °C
  - APT.line™ preheating chamber technology
  - Natural convection
  - Adjustable exhaust air flap
- Controller with timer function
  - 1 chrome-plated rack, incl. shelf supports
  - Class 2 integrated independent adjustable temperature safety device (DIN 12880) with visual alarm

## ORDERING INFORMATION


Interior volume [L]	Power supply - unit fuse	Plug*	Version	Model version	Art.-No.
Model ED-S 56					
62	120 V 1~ 60 Hz -6,3 A	NEMA 5-15	Standard	EDSo56UL-120V	9090-0015
	230 V 1~ 50/60 Hz -6,3 A	CEE 7/7	Standard	EDSo56-230V	9090-0014

## TECHNICAL DATA

Data		
Designation	EDSo56-230V	EDSo56UL-120V
Article Number	9090-0014	9090-0015
Option model	Standard	Standard
Performance Data Temperature		
Temperature range	+7 °C above ambient temperature to 250 °C	+7 °C above ambient temperature to 250 °C
Temperature uniformity at 150°C	3,4 ± K	3,4 ± K
Temperature fluctuation at 150°C	0.8 ± K	0.8 ± K
Heating up time to 150°C	60 min	60 min
Recovery time after door was opened for 30 s at 150°C	45 min	45 min
Electrical data		
Rated Voltage	230 V	120 V
Power frequency	50/60 Hz	60 Hz
Nominal power	1,05 kW	1,15 kW
Unit fuse	6,3 A	6,3 A
Phase (Nominal voltage)	1~	1~
Dimensions and weights		
Interior volume	62 L	62 L
Net weight of the unit (empty)	36 kg	36 kg
Load per rack	15 kg	15 kg
Permitted load	45 kg	45 kg
Wall clearance back	160 mm	160 mm
Wall clearance sidewise	100 mm	100 mm
Housing dimensions not incl. fittings and connections		
Width net	614 mm	614 mm
Height net	625 mm	625 mm
Depth net	577 mm	577 mm
Internal Dimensions		
Interior width	400 mm	400 mm
Interior height	440 mm	440 mm

Interior depth	350 mm	350 mm
Unit doors	1	1
<b>Environment-specific data</b>		
Energy consumption at 150°C	210 Wh/h	210 Wh/h
<b>Fixtures</b>		
Number of shelves (std./max.)	1/3	1/3

## DIMENSIONS Incl. fittings and connections [mm]



## ACCESSORIES

Designation	Description	ED-S 56	*	Art.-No.
Grid shelf	Incl. shelf supports			
	chrome plated	•	-	8012-1807
pH-neutral detergent	concentrated, for gentle remove of residual contaminants; 1 kg	•	-	8012-2250
Rubber pads	set anti-slip feet	•	-	8012-2030

## SERVICES

Designation	Description	*	Art.-No.
<b>Calibration services</b>			
Calibration certificate, temperature	Calibration of one (1) test temperature specified by the user in center of chamber, including certificate	-	DL30-0110
	Extension of calibration of one (1) additional test temperature specified by the user in the center of the usable space, including certificate	-	DL30-0102
Temperature measurement, 9 measuring points	Temperature measurement with 9 measuring points with a set value specified by the user, including certificate	-	DL30-0109
Temperature measurement, 27 measuring points	Temperature measurement with 27 measuring points with a set value specified by the user, including certificate	-	DL30-0127
<b>Installation services</b>			
Unit commissioning	Connect the unit to the customer-side connections (electricity, water, wastewater, gas), basic functions check, brief operating instructions. (excl.: unpacking, setup, controller instructions, programming, installation work)	-	DL10-0100
Unit instructions	Instruction regarding operating principle and basic functions of the unit, operation of the control electronics including programming	05	DL10-0500

Designation	Description	*	Art.-No.
<b>Maintenance contracts</b>			
BRONZE 3-year maintenance contract	Maintenance service as contractually agreed, visual inspection of mechanical and electrical components, check of control response, 20% discount on spare parts	-	DL20-0720
GOLD 3-year maintenance contract	Maintenance service as contractually agreed, visual inspection of mechanical and electrical components, check of control response, 20% discount on spare parts, testing of all key functions, replacement of wear parts, calibration of one test temperature specified by the user in the center of the usable space, including certificate	-	DL20-0915
SILVER 3-year maintenance contract	Maintenance service as contractually agreed, visual inspection of mechanical and electrical components, check of control response, 20% discount on spare parts, testing of all key functions, calibration of one test temperature specified by the user in the center of the usable space, without certificate	-	DL20-0815
<b>Maintenance services</b>			
Maintenance	One-off maintenance service in accordance with maintenance schedule. Visual inspection of mechanical and electrical components, testing of all key functions. Calibration of a test temperature specified by the user in center of usable space without certificate	-	DL20-0601
<b>Warranty service</b>			
1-year warranty extension	The warranty is extended by 1 year from the delivery date, wear parts are excluded	-	DL50-0010

## NOTES

- 05 Quoted prices do not include travel costs. Please refer to the chapter on BINDER Service for travel costs for your region. Quoted prices for services performed in Switzerland do not include a country-specific added fee (available on request).





# WolfLabs

**Pricing on any accessories shown can be found by keying the part number into the search box on our website.**

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

**[www.wolflabs.co.uk](http://www.wolflabs.co.uk)**

**Tel : 01759 301142**

**Fax : 01759 301143**

**[sales@wolflabs.co.uk](mailto:sales@wolflabs.co.uk)**

Please contact us if this literature doesn't answer all your questions.